



Product Service

CERTIFICATE

No. Z1A 15 10 12906 237

Holder of Certificate: DÜPERTHAL SICHERHEITSTECHNIK
GMBH & Co. KG



DÜPERTHAL
SICHERHEITSTECHNIK

Frankenstraße 3
63791 Karlstein
GERMANY

Factory(ies):

62099

Certification Mark:



Product:

Safety cabinets

Model(s):

ACID line, 20-HHBBTu-xyz(ss)
ACID line pure, 20-HHBBTu-xyz(ss)
ACID UTS line, 20-HHBBTu-xyz(ss)

Parameters:

Key code Art.-no.:

HH height, stand for numbers: 06, 13, 20
BB width, stand for numbers: 06, 11, 12
T depth, stand for numbers: 6, 7
u door closing technique, stand for numbers: 4, 7
x colouring, stand for numbers: 0, 4
yz interior, stand for numbers: 00 bis 99
ss Special sign, stand for numbers: L, R

Class of fire resistance: FWF 0

Thearticle number defines exactly the material an the size of the product. A full description of the specification is provided in the test Report.

Tested according to:

PPP 52125A:2011
DIN EN 14727:2006
AfPS GS 2014:01 PAK

The product meets the safety and health requirements of the German Product Safety Act section 20 to 22 ProdSG. The certification marks shown above can be affixed on the product. It is not permitted to alter the certification marks in any way. In addition the certificate holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. See also notes overleaf.

Test report no.:

713056799

Valid until:

2020-10-12

Date, 2015-10-20

(Horst Kristen)

Page 1 of 1

